



BROWLIFT BONE BRIDGE DRILL BIT

Warnings / Precautions	Before sterilization, carefully inspect the drill bit under a microscope or magnifying glass for any irregularities or eccentricities. Cold soak in glutaraldehyde, chlorine, or ammonium solutions, or dry heat sterilization is not recommended as damage to the instrument finish may occur.				
Limitations	Discard any drill bit that shows signs of damage or wear				
INSTRUCTIONS					
Point of Use	No particular requirements				
Containment and Transportation	It is recommended that instruments are reprocessed as soon as is practical following use. An available instrument tray (Ref: 37-47103) is recommended for sterilization and handling.				
Preparation for decontamination	Promptly and thoroughly rinse instruments with deionized water after each use.				
Cleaning: Automated	Remove instruments and equipment from any sterilization trays before placing into washer baskets. Orient devices following recommendations of washer/disinfector manufacturers. Use alkaline or neutral pH detergent recommended by washer/disinfector or detergent manufacturers. These products have been validated for effective cleaning using an automatic washer/disinfector cycle consisting of a minimum 44 minutes total time, including a pre-wash, main wash & rinse, and thermal rinse. The thermal rinse shall be at least 10 minutes long at a minimum temperature of 60°C.				
Cleaning: Manual	Soak in lukewarm*, mild* enzymatic detergent, and deionized water for a minimum of two minutes. Then clean ultrasonically in lukewarm* solution of mild* detergent and deionized water for at least 30 seconds. Rinse thoroughly with deionized water and wipe dry. *less than 43°C; pH 7.0 - 8.5 Note: When using an ultrasonic cleaner or a spray washing machine, follow the manufacturer's recommendations, particularly with regard to articulated instruments and positioning of instruments.				
Disinfection	No particular requirements				
Packaging	No particular requirements				
Sterilization (Temperatures are minimum required, times are minimum required)	Cycle:	Gravity	Pre-vac		
	Temperature: Time: Drying:	121°C 30 min 8 min vacuum	134°C 3 min 8 min vacuum		
	100% EtO Not validated			STERRAD Sterilization Not validated	
Maintenance, Inspection and Testing	No particular requirements. Discard any drill bit that shows signs of damage or wear				
Storage	No particular requirements				
Additional Information	None				

Note: The instructions provided above have been validated by the manufacturer as being CAPABLE of preparing the product for re-use. It remains the responsibility of the processor to ensure that the reprocessing as actually performed using equipment, materials and personnel in the reprocessing facility achieve the desired result. This normally requires validation and routine monitoring of the process.