



E/Z LOAD T-TUBE INSERTER

Warnings / Precautions	To prevent stains, use distilled or demineralized water, and use a neutral detergent, to reprocess these instruments		
Limitations	After cleaning and sterilization, verify functionality prior to re-use.		
INSTRUCTIONS			
Point of Use	This product is provided non-sterile and must be cleaned and sterilized before the first use and any reuse. Rinse instrument immediately after use in distilled or demineralized water		
Containment and Transportation	No particular requirements It is recommended that instruments are reprocessed as soon as is practical following use.		
Preparation for decontamination	Disassemble instruments as necessary for effective cleaning.		
Cleaning: Automated (Do NOT use ultrasonic washer)	Remove instruments and equipment from any sterilization trays before placing into washer baskets. Orient devices following recommendations of washer/disinfector manufacturers. Use alkaline or neutral pH detergent recommended by washer/disinfector or detergent manufacturers. These products have been validated for effective cleaning using an automatic washer/disinfector cycle consisting of a minimum 44 minutes total time, including a pre-wash, main wash & rinse, and thermal rinse. The thermal rinse shall be at least 10 minutes long at a minimum temperature of 60°C. Examine instruments for any staining or deterioration; remove from use as appropriate. Following cleaning, lightly lubricate instruments with movable parts. Use a lubricant intended for sterilizable instruments such as a water-soluble instrument milk. Do not use silicone spray.		
Cleaning: Manual	Soak in lukewarm (less than 43°C), mild (pH 7.0 - 8.5), enzymatic detergent and deionized water for a minimum of 2 minutes, then clean ultrasonically in a lukewarm (less than 43°C), mild (pH 7.0 - 8.5), detergent and deionized water for at least 30 seconds. Rinse thoroughly with deionized water immediately and dry and wipe dry with a lint-free cloth. Examine instruments for any staining or deterioration; remove from use as appropriate. Following cleaning, lightly lubricate instruments with movable parts. Use a lubricant intended for sterilizable instruments such as a water-soluble instrument milk. Do not use silicone spray. Note: When using an ultrasonic cleaner or a spray washing machine, follow the manufacturer's recommendations, particularly with regard to articulated instruments and positioning of instruments.		
Disinfection	No particular requirements		
Packaging	No particular requirements		
Sterilization (Temperatures are minimum required, times are minimum required)	Cycle:	Gravity or Pre-vac	
	Temperature:	121°C	
	Time:	30 min	
	Drying: 8 min or until visibly dry		
	EtO: Not validated		STERRAD Sterilization: Not validated
Maintenance, Inspection and Testing	Inspect components for any damage before and after each use. If damage is observed do not use the instrument until it is repaired or replaced. After cleaning and sterilization, verify functionality prior to re-use.		
Storage	No particular requirements		
Additional Information	None		

Note: The instructions provided above have been validated by the manufacturer as being CAPABLE of preparing the product for re-use. It remains the responsibility of the processor to ensure that the reprocessing as actually performed using equipment, materials and personnel in the reprocessing facility achieve the desired result. This normally requires validation and routine monitoring of the process.